



Republic of the Philippines  
 Department of Agriculture  
**BUREAU OF ANIMAL INDUSTRY**  
 5 Visayas Avenue, Barangay Vasra, Quezon City 1128

(+632)8528-2240 director@bai.gov.ph www.bai.gov.ph @bai.gov.ph

February 8, 2024

**SUPPLEMENTAL / BID BULLETIN**  
**No. 2024-002**

This Supplemental Bid Bulletin No. 2024-002 is being issued to advise prospective bidder/s on modifications and amendments on certain items in the Bidding Document and Terms of Reference, if applicable for the supply and delivery of various items:

FROM	TO
<b>IB -2024-006</b> <b>SUPPLY AND DELIVERY OF VARIOUS TUBES</b> <b>ABC - P2,785,000.00</b>	
1. 50ml Sterile Screw cap centrifuge tube <ul style="list-style-type: none"> <li>• Self-standing</li> <li>• conical bottom</li> </ul> 3. 1.5ml Sterile Screwcap centrifuge tubes with O ring  5. 0.2 mL PCR Tubes <ul style="list-style-type: none"> <li>• Natural color</li> <li>• Polypropylene</li> <li>• flat caps</li> <li>• DNase/RNase free</li> <li>• 1000pcs/pack</li> </ul> <b>Delivery Period:</b> 30-45 calendar days upon receipt of NTP	1. 50ml Sterile Screw cap centrifuge tube <ul style="list-style-type: none"> <li>• Self-standing <b>with conical bottom</b></li> <li>• <del>conical bottom</del></li> </ul> 3. 1.5ml – <b>2mL</b> Sterile Screwcap ( <b>external</b> ) centrifuge tubes with O ring  5. 0.2 mL PCR Tubes <ul style="list-style-type: none"> <li>• Natural color</li> <li>• Polypropylene</li> <li>• flat caps</li> <li>• DNase/RNase free</li> <li>• 1000pcs/pack</li> <li>• <b>Individual tubes</b></li> </ul> <b>Delivery Period:</b> <b>60</b> calendar days upon receipt of NTP
<b>IB -2024-007</b> <b>SUPPLY AND DELIVERY OF BIOANALYZER (AUTOMATED ELECTROPHORESIS)</b> <b>ABC - P3,900,000.00</b>	
Approximate Dimensions: 435mm x 510mm x 445mm (WxDxH)  <b>Inclusions:</b> 1. Laptop with automated electrophoresis system software  Inclusions: a. High Sensitivity D5000 ScreenTape, for 105 samples (2 pcs) b. High Sensitivity D5000 Reagents, for 105 samples (2 boxes) c. High Sensitivity Ladder, (includes 20 ul ladder) (2 boxes)	<b>Maximum</b> Dimensions: <b>800mm x 560mm x 880mm</b> (WxDxH)  <b>Inclusions:</b> 1. Laptop with automated electrophoresis system software <b>with compatible printer</b>  <b>4. compatible AVR</b>  Inclusions: a. High Sensitivity <b>cartridges</b> D5000 ScreenTape, for <b>minimum of</b> 105 samples (2 pcs) b. High Sensitivity <b>D5000</b> Reagents, for <b>minimum of</b> 105 samples (2 boxes) c. High Sensitivity D5000 Ladder, ( <b>minimum of</b> 20 ul ladder) (2 boxes)



"Our organization is certified according to ISO 9001"

<p>Installation and on-site training (training from a Technical Specialist)</p> <p>Certificate of Satisfactory Performance from previous deliveries</p> <p>Preventive maintenance and after-sales service within the warranty period</p>	<p>1. Installation and on-site training (training from a Technical Specialist)</p> <p><b>2. Training Certificate (updated) – Technical Specialist /Engineer)</b></p> <p>3. Certificate of Satisfactory Performance from previous deliveries <b>(minimum of 3)</b></p> <p>4. Preventive maintenance and after-sales service within the warranty period</p>
<p><b>IB -2024-008</b> <b>SUPPLY AND DELIVERY OF GENETIC ANALYZER</b> <b>ABC - P13,050,000.00</b></p>	
<p>Also included: 2KVA UPS with built- in AVR</p> <p>Installation and on-site training (training from a Technical Specialist)</p> <p>Certificate of Satisfactory Performance from previous deliveries</p> <p>Preventive maintenance and after-sales service within the warranty period</p>	<p>Also included: 2KVA UPS with built- in AVR</p> <p>1. Installation and on-site training (training from a Technical Specialist)</p> <p><b>2. Training Certificate (updated) – Technical Specialist /Engineer)</b></p> <p>3. Certificate of Satisfactory Performance from previous deliveries <b>(minimum of 3)</b></p> <p>4. Preventive maintenance and after-sales service within the warranty period</p>

Please be guided accordingly.

  
**RENE C. SANTIAGO, DVM MSc.**  
Chairperson  
Bids and Awards Committee

Received: \_\_\_\_\_  
Bidder/Representative  
Date: \_\_\_\_\_

