



Republic of the Philippines
Department of Agriculture
BUREAU OF ANIMAL INDUSTRY
5 Visayas Avenue, Barangay Vasra, Quezon City 1128

(+632)8528-2240 director@bai.gov.ph www.bai.gov.ph @bai.gov.ph

September 12, 2024

SUPPLEMENTAL / BID BULLETIN
No. 2024-018

This Supplemental/Bid Bulletin No. 2024-018 is issued by the BAI Bids and Awards Committee to all participating bidders to clarify, amend and/or modify certain provisions and requirements for the procurement of various projects.

All portions of the Bidding Documents affected by these amendments shall be made to conform the same.

Amendments/inclusions/clarifications made herein shall be considered as integral part of the Bidding Documents.

FROM	TO
IB -2024-068 SUPPLY AND DELIVERY OF REAL-TIME PCR MACHINE SYSTEM ABC – P9,626,000.00	
Specifications: Block Change: Fixed Display Type: Touch screen Format: 96-well plate Volume (Metric) Thermal Block Sample: 0.1 mL Dimension (W x D x H): 27 x 50 x 40 cm Features: Front-facing fully motorized block Touch screen interface Less frequent maintenance 10 GB on-board memory for storage of 2,000–5,000 runs USB interface Content and Storage: Computer	Specifications: Block Change: Fixed Display Type: Touch screen Format: 96-well plate Volume (Metric) Thermal Block Sample: 0.1 mL Maximum Dimension (W x D x H): 27 x 50 x 40 cm Features: Front-facing fully motorized block Touch screen interface Less frequent maintenance 10 GB on-board memory for storage of 2,000–5,000 runs USB interface Content and Storage: Computer
IB -2024-069 SUPPLY AND DELIVERY OF VORTEX MIXER AND MINI CENTRIFUGE ABC – P1,100,000.00	



"Our organization is certified
according to ISO 9001"

Masaganang Agrikultura, Maunlad na Ekonomiya



Republic of the Philippines
Department of Agriculture
BUREAU OF ANIMAL INDUSTRY
5 Visayas Avenue, Barangay Vasra, Quezon City 1128

(+632)8528-2240 director@bai.gov.ph www.bai.gov.ph @bai.gov.ph

VORTEX MIXER

Speed: 4,500 rpm

Certifications: CE, CB Scheme, ROHS3, WEEE and SGS

Warranty: 5 year

MINI CENTRIFUGE

Certifications: UL (Canada and US), ROHS3, CE, WEEE, and C-Tick (RCM)

Warranty: 5 years

Time of delivery: 60 days

VORTEX MIXER

Minimum Speed: 4,500 rpm

Certifications: CE/CB Scheme/ROHS3/WEEE/SGS

Warranty: 3-5 year **on parts and services**

MINI CENTRIFUGE

Certifications: UL (Canada and US)/ROHS3/CE/WEEE/C-Tick (RCM)

Warranty: 3-5 years **on parts and services**

Time of delivery: 60 **calendar** days **upon receipt of NTP**

IB -2024-070

SUPPLY OF LABOR AND MATERIALS FOR THE COMPREHENSIVE PEST CONTROL AND MAINTENANCE SERVICE FOR ONE (1) YEAR
ABC – P1,390,000.00

TERMITE CONTROL

-Dusting Method

Minimum Requirements:

- Must be a member with good standing of the Pest Control Association of the Philippines (PCAP) or Philippine Federation of Pest Management Operators Association (PFPMOA) or Pest Exterminators Association of the Philippines (PEAP)
- Must have the latest license of Pest Control Operator by Fertilizer and Pesticide Authority (FPA) or by Food and Drug Administration (FDA)

TERMITE CONTROL

- **Baiting Method**

Minimum Requirements:

- Must be a member with good standing of the Pest Control Association of the Philippines (PCAP) or Philippine Federation of Pest Management Operators Association (PFPMOA) or Pest Exterminators Association of the Philippines (PEAP) or **FDA Recognized Pest Control Operators Association**
- Must have the latest **valid** license of Pest Control Operator by Fertilizer and Pesticide Authority (FPA) or by Food and Drug Administration (FDA)

Frequency of treatment: quarterly for GPC (General Pest Control)

To be submitted as part of the bidding docs:


- *Certificate of site inspection**
- *Pre-assessment or inspection report**

Payment: progressive billing
1st payment: 60% completion



"Our organization is certified according to ISO 9001"

Masaganang Agrikultura, Maunlad na Ekonomiya

<p>Checklist of Technical and Financial Documents</p> <p>5. Conformity with the Technical Specifications</p> <ul style="list-style-type: none"> Conformity with the Technical Specifications which may include Production/delivery schedule Manpower requirements After- sales/parts, if applicable 	<p>Succeeding payment: quarterly, 10%</p> <p>Delivery Period: 1st treatment, within 30 calendar days upon receipt of NTP</p> <p>Checklist of Technical and Financial Documents</p> <p>5. Conformity with the Technical Specifications</p> <ul style="list-style-type: none"> Conformity with the Technical Specifications which must include Production/delivery schedule Manpower requirements After- sales/parts, if applicable
<p>IB -2024-071 PROCUREMENT OF SERVICE FOR THE PRODUCTION OF AUDIO AND VIDEOS FOR AFRICAN SWINE FEVER (ASF) (SECOND POSTING) ABC – P4,000,000.00</p>	
<p>PLEASE SEE ATTACHED TERMS OF REFERENCE AND CONDITIONS</p>	
<p>IB -2024-072 SUPPLY AND DELIVERY OF VARIOUS LABORATORY EQUIPMENT ABC – P2,898,400.00</p>	
<p>LOT I: WATER DISTILLER</p> <ul style="list-style-type: none"> With external dispenser (detachable and ergonomically shaped, rotatable & adjustable) with 5" touchscreen Comes with 1-unit dispenser, level sensor, circulation pump <p>Pure water specifications (ASTM Type 2)</p> <ul style="list-style-type: none"> Delivery flow rate: 6L/h; Resistivity: > 10MOhm-cm <p>Ultrapure Water Specifications (ASTM Type 1)</p>	<p>LOT I: WATER DISTILLER Reference picture for 5" touchscreen:</p>  <ul style="list-style-type: none"> With accessories of 1-unit dispenser, level sensor, circulation pump, sterile overflow, and vent filter <p>Pure water specifications (ASTM Type 2)</p> <ul style="list-style-type: none"> Delivery flow rate: Minimum 6L/h; Resistivity: > 10MOhm-cm <p>Ultrapure Water Specifications (ASTM Type 1)</p> <ul style="list-style-type: none"> TOC: ≤ 2 ppb (with UV lamp); ≤ 10 ppb (w/o UV lamp)



Republic of the Philippines
Department of Agriculture
BUREAU OF ANIMAL INDUSTRY

5 Visayas Avenue, Barangay Vasra, Quezon City 1128

(+632)8528-2240 director@bai.gov.ph www.bai.gov.ph @bai.gov.ph

<ul style="list-style-type: none"> • TOC: ≤ 2 ppb (with UV lamp); ≤ 10 ppb (w/o UV lamp) • Bacteria: < 0.01 CFU/ml (with sterile capsule) < 0.005 CFU/ml (with ultrafiltration module) <p>Feed Water Requirements</p> <ul style="list-style-type: none"> • Feed Water Quality: Potable Tap Water; Inlet Pressure: 1-6 bar • Temperature: 5 - 35°C • Conductivity at 25 °: < 2000 μS/cm; Total Organic Carbon (TOC): < 1 ppm <p>Technical Data</p> <ul style="list-style-type: none"> • Dimensions, mm: 410Wx465Dx765H (with tank below) <p>Delivery Period: 30 calendar days</p>	<ul style="list-style-type: none"> • Bacteria: < 0.005 CFU/ml (with ultrafiltration module) <p>Feed Water Requirements</p> <ul style="list-style-type: none"> • Feed Water Quality: Potable Tap Water; • Inlet Pressure: 1-6 bar • Temperature: 5-35°C • Conductivity at 25 °: < 2000 μS/cm; Total Organic Carbon (TOC): < 1 ppm <p>Technical Data</p> <ul style="list-style-type: none"> • Dimensions, mm: Maximum 410Wx465Dx765H (with tank below) <p>With inclusion of:</p> <ul style="list-style-type: none"> • In-line ultrafilter with a recommended change interval of least 3 years or up to 20,000 liters. • Integrated pre-treatment module for a direct connection to tap water, and to guarantee optimum pure water quality. • Internal recirculation (including tank) up to the dispensing point to avoid microbial contamination and ensure water quality. • With precise and variable volume dispensing <p>Delivery Period: within 60 calendar days upon receipt of NTP</p>
<p>LOT II: FLUOROMETER</p> <ul style="list-style-type: none"> • 7.0 inch, color touch screen • Input voltage: 24 VDC 2A • Dimensions (WxDxH): 16.1x28.7x7.6cm / 6.3 x 11.3 x 3.0 inc • Weight: 2.0kg/4.4lbs <p>Delivery Period: 30 calendar days</p>	<p>LOT II: FLUOROMETER</p> <ul style="list-style-type: none"> • 7.0 inch (maximum) color touch screen • Input voltage: 24 VDC 2A and operating power 240V • Maximum Dimensions (WxDxH): 6.3 x 11.3 x 3.0 inc • Maximum Weight: 2.0kg/4.4lbs <p>Delivery Period: 60 calendar days upon receipt of NTP</p>



"Our organization is certified according to ISO 9001"

Masaganang Agrikultura, Maunlad na Ekonomiya

**LOT III: MICROVOLUME UV-VIS
SPECTROPHOTOMETER**

- Detector: CCD, 2048 pixels
- Dimensions/weight: 220x280x220mm/3kg

Delivery Period: 30 calendar days

**LOT III: MICROVOLUME UV-VIS
SPECTROPHOTOMETER**

Detector: CMOS 4096 pixels

Maximum Dimensions/weight:
220x280x220mm/3kg

Cuvette Specifications:

- 3 pcs

With the inclusion of:

- fluorometer accessory
- Detectors Solid state silicon sensor
- Excitation Filters Red: 640nm shortpass (wavelengths up to 640nm)
- Blue: 495nm short-pass (wavelengths up to 495nm)
- Emission Filters Red: 660-720nm Blue: 510-580nm
- Calibration Type Single-point (blank and standard)

Delivery period: **Within 60 calendar days upon receipt of NTP**

Please be guided accordingly.


CHERYL ROSE CAYAD-AN, DVM, MDM
Chairperson
Bids and Awards Committee