



May 10, 2016

**SUPPLEMENTAL / BID BULLETIN**  
**No. 2016-009**

This Supplemental Bid Bulletin No. 2016-009 is being issued to advise prospective bidders to modify and amend certain item in the Bidding Document and Invitation to Apply and Bid as follows:

Query	Answer
<p><b>BAI Internet Connectivity</b></p> <p>a. Is this new connection or renewal?</p> <p>b. Can we be allowed to have sub-contract?</p> <p>c. Are we going to submit document for the international gateways?</p> <p>d. What point of destinations to use to leverage with the international gateways?</p>	<p><i>A new connection</i></p> <p><i>Yes, provided that the main company will be liable to any circumstances encountered on the implementation of the project until its warranty period</i></p> <p><i>Yes, prospective bidder must submit any document to support the international gateways.</i></p> <p><i>The destinations are U.S., Middle East, Dubai, South Korea, Japan, Singapore and Riyadh</i></p>
<p>Supply and delivery of 1 unit laboratory steam sterilizer</p> <p>a. What do you mean by delayed start function with real time allows the users to find their samples sterilized when they arrive to the laboratory in the morning so they do not loose time waiting for the sterilization of their samples?</p> <p>b. Can we change the line --- the microprocessor control systems have a comprehensive self-diagnostic system for providing information regarding system malfunctions?</p> <p>c. Can we offer bigger chamber volume?</p> <p>d. The preset programs are for what kind of 2 liquids and 3 solids?</p> <p>e. Can we remove the timer for free programs: 1-300 minutes?</p> <p>f. What is the real power consumption?</p> <p>g. Can the delivery period be increased?</p>	<p><i>It was agreed to remove this line</i></p> <p><i>It was agreed to change to <b>it must have comprehensive self-diagnostic system.</b></i></p> <p><i>The chamber volume will be <b>minimum of 40 liters</b></i></p> <p><i>The two (2) liquids are <b>culture media and reagents</b> while the three (3) solids are <b>glass wares, syringes and autoclavable canisters</b></i></p> <p><i>It was agreed to remove this line.</i></p> <p><i>The power consumption is <b>2,500W</b> and not 250 00W.</i></p> <p><i>The delivery period will be <b>from 30 to 90 days.</b> You can also deliver the goods within the specified period.</i></p>





**Republic of the Philippines**  
 Department of Agriculture  
**BUREAU OF ANIMAL INDUSTRY**  
 Visayas Avenue, Diiman, Quezon City

h. What do mean by actual demonstration of the product? Does this mean all the bidder will have to demonstrate the product before the opening of bids or upon delivery?	<i>The actual demonstration will be done <b>upon the delivery of the product.</b></i>
i. What is the warranty period?	<i>The warranty period is <b>minimum of one (1) year.</b></i>
Supply and delivery of agricultural accessories How many blades are needed for the disc harrow	   

Reminder:

Guidelines for the submission of some documents:

- a. Expiry of Tax Clearance must be six (6) months before the opening of bids;
- b. If tax clearance is for renewal, submit the previous copy including the claim stub;
- c. Computation of the Net Financial Contracting Capacity (NFCC) must use the latest Audited Financial Statements received and stamped by the BIR; and
- d. Indicate on the Bid Securing Declaration the ITB # of the item to be bid

Please be guided accordingly.

Approved by:

**MARIA V. ABENES, RMT, DVM, MM**  
 Chair, Bids and Awards Committee

Received by the Bidder/Representative:

Date: \_\_\_\_\_

