


July 9, 2020

## SUPPLEMENTAL / BID BULLETIN No. 2020-06

This Supplemental Bid Bulletin No. 2020-06 is being issued to advise prospective bidder/s on modifications and amendments on certain items in the Bidding Document, Invitation to Apply and Bid and Terms of Reference, if applicable for the infrastructure projects:

ITB/TOR Specifications	Agreements during Pre-Bid Conference
ITB -2020-018  Rebidding for the supply and delivery of one (1) lot Biosafety Cabinet Class II Type C1	
ABC - P 1,200,000.00	
Specifications:  • inclusive of HEPA filter for user safety with 3" metal frame supply HEPA Filter 99.99% @0.3 microns	Corrections:  • may offer HEPA (High efficiency particulate air filter or ULPA (Ultra-low particulate air filter)
*Inclusive of installation fees and services Third Party validation (one-time)	* Filters should be readily available in the market  *Must attach Exclusive / Authorized Distributorship Certificate
	*Must submit list of factory trained/service engineers with training certificate to serve as protection for preventive maintenance  *Inclusive of installation fees and services with performance verification after installation and commissioning
Please be guided accordingly.	Received: Bidder/Representative





BAC Chair- Goods & Consultancy Services