



May 19, 2022

Date

**REQUEST FOR QUOTATION
 SMALL VALUE PROCUREMENT**

INSTRUCTION:

- Bidders are required to read the instructions and fill all the blanks properly, Prospective Bidders/Suppliers are required to use this official canvass form in accomplishing their bid proposals/quotation.
- Any specifications other than those required/stated in this form shall not be considered in their evaluation of bid.
- Quotation must be the lowest price(s), taxes included for the item(s) services listed hereunder including delivery charges.
- Always indicate the brand name of the offered items (s) or product.
- Price quotation(s) submitted shall be valid for ONE HUNDRED TWENTY DAYS (120) days reckoned from the deadline for submission of quotations.
- Terms of payment-Thirty (30) days from delivery of items(s)
- Only sealed quotations shall be considered by BAC. The Bidders/Suppliers or Authorized Representative shall indicate on the envelope the RFQ Number and the company name of the bidder/supplier.**
- Awarding shall be done by **LOT /CATEGORY** only.
- Delivery and/or Installation Period: 60 working days upon receipt of Notice to Proceed

DEADLINE FOR SUBMISSION OF BIDS

May 24, 2022

4:00 PM

ASF 2022-

Date

Time

PR No. 04-090

End-user:

MITZI LUISA C. MORALES, RMT

ABC: 250,000.00

Qty.	Unit	Item/Description	ABC (PhP)	Offered brand	Unit price	Total cost
1	unit	Cell Culture Incubator	250,000.00			
		External Dimensions (width x Depth x Height)				
		500mm (19.7") x 500mm (19.7") x 655mm (25.8")				
		Internal Dimensions (width x Depth x Height)				
		345mm (13.6") x 375mm (14.8") x 388mm (15.3")				
		Volume 50L (1.8 ft^3)				
		Material				
		Main body: 1.2 mm (0.05") 18-gauge electro-galvanized steel with white oven-baked epoxy-polyester				
		Work zone: abrasion and corrosion-resistant 304 grade, 0.9 mm stainless steel with BA finish				
		Shelves				
		Material: 1.5 mm, grade 304 stainless steel shelves				
		3 pcs standard number of shelves				
		11kg/shelf (24.3 lbs/shelf) maximum load per shelf				
		Direct Heat and Air Jacket				
		Direct heat achieves precise heating through the 8 heaters located in 3 zones for rapid temperature recovery and improved temperature uniformity				
		Air jacket provides insulation of the chamber from ambient temperature fluctuations				
		temperature control range: ambient +7 to 60 deg C				
		temperature uniformity: <+/-0.5 deg C				
		temperature accuracy: <+/-0.5 deg C				
		temperature recovery (after 30 seconds door opening): < /= 5				
		Forced Convection System				
		Blower fan circulates filtered air in the chamber thereby accelerating the humidifying process and improving CO2, humidity and temperature uniformity				
		Humidity range is 85%-90% using humidity fan				
		Heat Resistant Infrared (IR) C)2 Sensor				
		Drift-free sensor that enhances CO2 control and uniformity without being affected by temperature and pressure				
		Has auto-zeroing feature which eliminates the need for calibration				
		% CO2 control range: 0-19.5%				
		% CO2 accuracy: +/- 1%				
		% CO2 recovery time (after 30 seconds door opening): < /= 6 minutes				
		Moist Heat Decontamination Cycle				
		Effectively deactivates normally-resistant fungi, bacterial spore and vegetative cells validated independently and by the Health Protection Agency (HPA) in UK				
		0.2um Gas Inlet Filter				
		Removes impurities and contaminants from gas before it enters the Four-line LCD display				
		Intelligent data and event logger records all incubator parameters for 2 MB build-in flash memory				
		guarantees long-term storage of data				
		With one touch decontamination button				
		Has user-configurable audible and visual alarms equipped				
		Electrical Requirements				
		Power rating: 220-240 VAC, 50/60 Hz, 1Ø				
		Normal Power at 37 deg C: 40.9 W@				
		Maximum Power Consumption: 598.8 WQ				
		Must Include				
		Automatic Voltage Regulator (AVR)				
		CO2 Regulator				
		Period of delivery: 60 days				
		Place of delivery: ADDRL- Molecular Laboratory				

CANVASSER'S CERTIFICATION

This is to certify that I have full knowledge, authority and responsibility in distributing and/or collecting the Request for Quotation in accordance to the guidelines in securing prices for the Bureau of Animal Industry.

(Signature Over Printed Name)

Authorized Canvasser

The BAI Bids and Awards Committee (BAC)
Visayas Avaneue, Diliman, Quezon City

Sir/Madam:

In connection with the above request, I/We hereby submit our quotation indicated above. I/We have carefully read and fully understand the minimum requirements and agree to furnish and/or effect delivery in conformity with specifications any or all said articles/services described above within _____ working days from receipt of Purchase Order.

Signature Over Printed Name

Name of Company

Tax Identification Number

Telephone Number(s)

Address

(To be filled by Supplier)
Date accomplished

Mandatory Requirements for submission:

- ✓ Mayor's Permit (Updated)
- ✓ PhilGEPS Registration/Certificate (Updated)
- ✓ Income Tax Return (for ABCs above P500k)
- ✓ Omnibus Sworn Statement (For ABCs above P50K) (Unnotarized OSS may be submitted provided that the notarized OSS shall be submitted prior to issuance of Notice of Award.