



August 22, 2017

SUPPLEMENTAL / BID BULLETIN
No. 2017-007

This Supplemental Bid Bulletin No. 2017-007 is being issued to advise prospective bidders to modify and amend certain item in the Bidding Document and Invitation to Apply and Bid for the following goods:


ITB #	Original Specification	Agreed Specification
Batch 1 2017- 048	Supply and delivery of 1 system Inductive Coupled Plasma Optical Emission Spectrophotometer Pixel size: at least 15µm Number of pixels: at least 2 mega pixels Software: <ul style="list-style-type: none"> • Industry standard MS Windows Accessories: <ul style="list-style-type: none"> • Hydride generator (if applicable) • UPS 10KVA, 200-240 VAC, 50/60Hz (with conversion from 3 Phase to single Phase output) Others: <ul style="list-style-type: none"> • Training for 3-4 persons with principal supplier inclusive of air fare, hotel and accommodation 	At most 30µm Retained Software: <ul style="list-style-type: none"> • Industry standard MS Windows or equivalent Accessories: <ul style="list-style-type: none"> • Hydride generator • UPS 10KVA, 200-240 VAC, 50/60Hz Others: <ul style="list-style-type: none"> • Training for 3 persons at manufacturer's site or training center / site • Must have a trained engineer with valid certificate (bring the original copy of the certificate during the opening of bids)
2017- 049	Supply and delivery of 2 units Ultrasound Scanner with Probes Multi- frequency Linear Rectal Probe Multi - frequency Linear Probe	Linear and not curvilinear Frequency: 5.0/7.5/10.0 MHz Width : 50mm Linear and not curvilinear Frequency: 1.5/2.0/3.0/3.5 MHz. Width: 180 mm

	Accessories (Mitsubishi) printer	Any brand of printer The accessories include an offset probe guide for each linear probe
--	----------------------------------	---

The Opening of Bids will be rescheduled from September 1, 2017 to **September 5, 2017**, same time and venue.

Please be guided accordingly.

Approved by:


MAGDALENA S. CRUZ, DVM
Chair, Bids and Awards Committee

Received by the Bidder/Representative: _____
Date: _____