



Republic of the Philippines
Department of Agriculture
BUREAU OF ANIMAL INDUSTRY
5 Visayas Avenue, Barangay Vasra, Quezon City 1128

(+632)8528-2240 director@bai.gov.ph www.bai.gov.ph @bai.gov.ph

December 20, 2024

SUPPLEMENTAL/BID BULLETIN
No. CEFA-SB-2024-028

This Supplemental/Bid Bulletin No. CEFA-SB-2024-028 is issued by the BAI Special Bids and Awards Committee (SBAC) for the Cold Examination Facility for Agriculture (CEFA) to all participating bidders to clarify, amend and/or modify certain provisions and requirements for the various procurement projects.

All portions of the Bidding Documents affected by these amendments shall be made to conform the same. Amendments/inclusions/clarifications/revisions made herein shall be considered as integral part of the Bidding Documents.

Technical and Scientific Equipment	LOT NO.	AMENDMENTS/MODIFICATIONS		
		FROM	TO	SPECIFIC SECTION OF THE BIDDING DOCUMENT
IB NO. CEFA-SB-2024-031 SUPPLY, DELIVERY AND INSTALLATION OF TECHNICAL AND SCIENTIFIC EQUIPMENT FOR CEFA SITE BULACAN – BATCH 5 (8 LOTS)				
Double Beam, Dual Atomic Absorption Spectrophotometer (AAS), Flame and Graphite Furnace with Autosampler	1	CEFA PR-SB-2024-085	CEFA-SBAC 2024-085	Section VI. Schedule of Requirements and Section VII. Technical Specifications
		Bandwidth: Must be settable to 0.2, 0.7, 1.30, and 2.0 L nm (4-step automatic switching)	Bandwidth: Must be settable to 0.2, 0.7, 1.0 , and 2.0 L nm (4-step automatic switching)	Section VII. Technical Specifications
Microwave Digester	3	CEFA PR-SB-2024-087	CEFA-SBAC 2024-087	Section VI. Schedule of Requirements and Section VII. Technical Specifications
		Top-loading microwave made of stainless steel with electrically-controlled swing top safety swiveling lid and circular oven for homogenous microwave distribution	Top-loading microwave made of stainless steel with electrically-controlled swing top safety swiveling lid and circular oven for homogenous microwave distribution, or equivalent	Section VII. Technical Specifications



"Our organization is certified according to ISO 9001"




Republic of the Philippines
Department of Agriculture
BUREAU OF ANIMAL INDUSTRY
5 Visayas Avenue, Barangay Vasra, Quezon City 1128

(+632)8528-2240 director@bai.gov.ph www.bai.gov.ph @bai.gov.ph

Technical and Scientific Equipment	LOT NO.	AMENDMENTS/MODIFICATIONS		
		FROM	TO	SPECIFIC SECTION OF THE BIDDING DOCUMENT
IB NO. CEFA-SB-2024-031 SUPPLY, DELIVERY AND INSTALLATION OF TECHNICAL AND SCIENTIFIC EQUIPMENT FOR CEFA SITE BULACAN – BATCH 5 (8 LOTS)				
Rotary Evaporator with Chiller	4	CEFA PR-SB-2024-088	CEFA-SBAC 2024-088	Section VI. Schedule of Requirements and Section VII. Technical Specifications
		Imersion ange adjustable to 7-stages enables the use of different flask size, adjustable for individual application	Immersion angle adjustable to 7-stages enables the use of different flask size, adjustable for individual application	Section VII. Technical Specifications

Please be guided accordingly.


MARIE BERNADETTE T. BONO, DVM
Vice-Chairperson, SBAC for CEFA

Received:

Bidder/Authorized Representative
Date: _____



"Our organization is certified according to ISO 9001"