



September 18, 2017


**SUPPLEMENTAL / BID BULLETIN**  
**No. 2017-011**

This Supplemental Bid Bulletin No. 2017-011 is being issued to advise prospective bidders to modify and amend certain item in the Bidding Document and Invitation to Apply and Bid for the following projects:

Original	Agreed / Recommended
<b>ITB #2017-056</b> Supply and delivery of Classical Swine Fever (Antibody & Antigen) test kits	Include: BAI registration
<b>ITB #2017-057</b> Supply and delivery of 1 unit Vacuum Oven - Appropriate control of standard components i.e temperature, timer and pressure - Shelves On-site calibration (ISO-17025)	Temp. range: 10 – 220°C Capacity: 100 – 127 liters Pressure: 5 – 1,100 mbar 2 – 3 shelves (layers), movable / adjustable Will be done by 3 <sup>rd</sup> Party Calibration for ISO-17025
Delivery period: 15 working days	Delivery: within 30-60 calendar days
<b>ITB #2017-064</b> Supply and delivery of 1 unit Horizontal <u>Rectangular</u> High Pressure Steam Sterilizer / Autoclave Chamber size: Diameter not less than 750mm; Length: not less than 1100mm Chamber volume capacity: not less than 485 liters	<b>Supply and delivery of 1 unit Horizontal High Pressure Steam Sterilizer / Autoclave</b> Chamber size: Width: 600mm; Height: 600mm; Depth: 1200mm Chamber volume capacity: not less than <b>425</b> liters
<b>ITB # 2017-058</b> Supply and delivery of Cold Handling Cabinet Delivery period: 15 working days	Delivery period: within 60-90 days (approved by BAC per formal inquiry of a supplier)

Please be guided accordingly.

Approved by:

  
**LAARNI Z. CABANTAC, DVM**  
 Chair, Bids and Awards Committee

Received by the Bidder/Representative: \_\_\_\_\_  
 Date: \_\_\_\_\_