



Republic of the Philippines
Department of Agriculture
BUREAU OF ANIMAL INDUSTRY
Visayas Avenue, Diiman, Quezon City

May 4, 2015

SUPPLEMENTAL / BID BULLETIN
No. 2015-003

This Supplemental Bid Bulletin No. 2015-003 is being issued to advise prospective bidders to modify and amend certain item in the Bidding Document and Invitation to Apply and Bid as follows:

Particulars	Agreed
Delivery period	The company can indicate their own delivery period. What is stated in the Invitation to Bid is the minimum number of days to delivery.
ITB #026-15 Supply & delivery of Lab. Eqpt. (Protein Digestor &	<ul style="list-style-type: none">- Digestion tube is 250ml, straight- Exhaust for stainless steel digestor- 1 can Kjel tabs – original packing (1000 pcs/can)- Stainless steel tube rack
Autoclave	<ul style="list-style-type: none">*External material can be stainless steel or powder coating*Wattage is maximum of 6,000*Capacity is minimum of 100L*Pressure: min 1.2kg/cm³*Temp. range: min. 121°C*Include appropriate/compatible AVR
ITB #027-15 - Spectrophotomer	Any equivalent software
ITB #028-15 - Supply and delivery of Lab. Eqpt. (Biosafety Cabinets	For Lot 2 – Internal size: min. 1800Wx600Dx660H mm Filter: can be ULPA or HEPA On-site validation must be done by a certified engineers
ITB #029-15 - Room Misting Device Sanosil disinfectant solutions	*Remove the word Sanosil and change to appropriate solution

Please be guided accordingly.

Approved by:


MAGDALENA S. CRUZ, DVM
Chair, Bids and Awards Committee

Received by the Bidder/Representative:



Mano
May 6, 2015